

# M9

# 9.0mm TO-Can

## STANDARD FEATURES

- 14.0 or 6.5mm pin length
- AR coated window
- AIN carrier, AuSn bonding
- Hermetically sealed package
- Plated gold over nickel pins



- Raman spectroscopy
- Laser therapy
- Laser pumping
- Medical
- Defense



## **SPECIFICATIONS**

Parameter	Unit	Min	Тур	Мах
Wavelength <sup>1</sup>	nm	-	785	-
Operating Power	mW	-	150	-
Operating Current	mA	-	170	210
Operating Voltage	V	-	-	2.2
Threshold	mA	-	35	55
Slope Efficiency	W/A	0.9	-	-
Vertical Far Field @ FWHM	deg. FWHM	-	25	-
Horizontal Far Field @ FWHM	deg. FWHM	-	8	-
Operational Temp <sup>2</sup>	°C	-20	25	50
Storage Temp	°C	-40	-	80
Lifetime (based at 25°C, lop, CW) <sup>3</sup>	hour	10,000	-	-

### **NOTES**

- 1) Wavelength options at ±3, 5 and 10nm.
- 2) Lifetime is based on operation at 25°C within a hermetically sealed package. Devices must be passively or actively cooled in accordance with the provided specifications. Failure to comply with heatsinking requirements may result in device failure.
- 3) Lifetime is quoted on accelerated CW testing.
- 4) See mechanical drawing (RDW 840015) for dimensions.



		1st digit
	R	SM PD, cathode ground
S SM P-up for sub		SM P-up for submounts

# CUSTOMIZE YOUR PACKAGE

Choose the last 3 digits of the part number to create your custom package

RPM9-785-0150-XXX (λ) (POWER) (123)

		2nd digit
	3	Wavelength tolerance ±3nm
5 Wavelength tol		Wavelength tolerance ±5nm
	9	Wavelength tolerance ±10nm

		3rd digit	
0	SM chip		

### OPERATING CONSIDERATIONS

Operating the diode laser outside of its maximum ratings may present a safety hazard or cause a device failure. Additionally, CW diode lasers may be damaged by excessive drive current or switching transients. When using a power supply with the component, it must be used within the specified parameters. DO NOT exceed the maximum peak optical power. Before turning the power supply on, connect the component to the power supply and ensure the output voltage value is zero. After the component has been successfully connected, increase the current slowly and monitor both the output power and drive current. Device degradation accelerates with increased temperature; therefore, careful attention to minimize the case temperature is advised. A proper heat-sink for the diode laser on a thermal radiator will greatly enhance laser life.

The primary cause of diode failure is unexpected electrostatic discharge. To help prevent device failures, be sure to handle devices with extreme care. The user should always wear an ESD wrist strap, ground all applicable work surfaces and follow anti-static techniques when handling diode lasers.

#### FDA 21 CFR 1040.10


All devices are manufactured, tested and labeled in compliance with FDA 21 CFR 1040.10 regulations, as applicable under the Radiation Control for Health and Safety Act of 1968. For smaller devices, the appropriate compliance labeling may be affixed to the shipping containter. All products comply with 21 CFR Chapter 1, Subchapter J.

Caution: Laser light emitted from a diode may be harmful to the human eye. Avoid looking directly into the diode laser aperture when the device is in operation. Note: The use of optical instruments with this product will increase eye hazard.

### CONTACT US

RPMC Lasers, Inc. 8495 Veterans Memorial Pkwy O'Fallon MO 63366 USA

MKT RPM9000002 REV A



WARNING! Invisible laser radiation is emitted from devices as

